


Smooth Silicone Coated Lancets!

Bayer's MICROLET®2 lancing device is used for obtaining capillary blood samples from the fingertip.



IMPORTANT: Do not use the same MICROLET2 lancing device on more than one person.

DIRECTIONS FOR USE:

 Prepare the puncture site. Wash your hands with soap and warm water. This helps increase blood flow to the fingertips. Rinse and dry thoroughly.

2. Remove the endcap.




In one hand, hold the MICROLET2 lancing device with your thumb on the grip indent as shown above. In your other hand, hold the endcap dial and gently snap off the endcap from top to bottom.

3. Hold a MICROLET lancet and ro-




Firmly push the lancet straight into the lancet holder. After you hear or feel the "click", continue pushing until the lancet comes to \subseteq


a complete stop. Twist off the protective cap and save for disposal of used lancet.








NOTE: If you need to 5. Select any of the five puncture depth settings by rotating the dial on the endcap to a smaller drop for a shallower puncture or larger drop for a deeper puncture. Adjusting the amount of pressure applied to the puncture site will also affect puncture depth.





re-cock the lancing ₹CLICK3 device, grip, pull or " push until device clicks as shown, and release.





7. To help form the blood drop, wait a few seconds. then stroke vourl hand and finger toward the puncture site. Do not squeeze around the \mid puncture site.

- 8. Test immediately after the blood drop has formed.
- 9. Wipe the puncture site with a clean, dry tissue and **maintain pressure** on the site until the bleeding stops.
- 10. To remove the used lancet, carefully remove the endcap by holding the endcap dial and snapping the endcap off.

Place the protective cap that you saved when you loaded the lancet on a flat surface with the Bayer logo face down. With the used lancet still in the


MICROLET2 lancing device, **push** the into the **center** of the cap.

Point the device toward a container for sharp objects (an \setminus

empty plastic bottle with a lid will work.) Label the container properly and keep out of the reach of children.

blue **release button** firmly pull the blue **cocking** handle (b) past resistance and the lancet will fall into the container.

Replace the end-



pose of the used lancet properly to prevent injury or contamination to others. Use a new MICROLET lancet each time you test 2. After the lancet is used, it is no longer sterile and has the potential to spread infection. All products or objects which come in contact with human blood, even after cleaning, should be handled as if capable of transmitting viral diseases.

Potential Biohazard: Always dis-

Clean your MICROLET2 lancing **device** after every use with a solution of water and mild detergent **or** a solution of nine parts water and one part bleach. Slightly moisten a tissue with the solution. Important: Avoid excess moisture. Remove the endcap. Clean the exterior of the device and interior and exterior of the endcap. Wipe dry with a tissue before use. **Replace** the endcap.

To contact Customer Service

If you need assistance or have guestions about blood glucose testing, please contact Bayer Diabetes Care Helpline at 1-800-268-7200.

STERILE R = Sterile

(2) = Do Not Reuse

Lancing device made in U.S.A.

Bayer, the Bayer Cross and MICROLET are trademarks of Bayer AG, used under license by Bayer HealthCare LLC and Bayer Inc.



(BANER) Bayer HealthCare

Manufactured by: Bayer HealthCare LLC Subsidiary of Bayer Corporation Mishawaka, IN 46544 U.S.A.

Distributed by: Bayer Inc. Toronto, Ontario M9W 1G6 Canada

Bayer

50241923 Rev. 5/08